

## Perelman School of Medicine Smilow Center for Translational Research Infectious Waste Disposal Guide\*

\* For research laboratories located in the Smilow Center for Translational Research and the South Tower ONLY. Procedures may vary in the hospital, clinical areas or other University spaces. For more information on infectious waste consult the University's Biological Safety Manual ([http://www.ehrs.upenn.edu/programs/bio/bio\\_manual.html](http://www.ehrs.upenn.edu/programs/bio/bio_manual.html)), or call EHRS at 215- 898-4453.

Container Type	Glassware/ Plasticware Waste Cardboard Box	Infectious Waste Sharps Containers	Infectious Waste Autoclave Bags	Infectious Waste Liquid Containers	Infectious Waste Bins located in Autoclave Room
<b>Contents</b>	<p>Uncontaminated laboratory glassware and plasticware. Chemical bottles must be triple rinsed and label defaced.</p> <p>Line cardboard glassware boxes with clear, heavy plastic bags.</p> <p><b>Do not use biohazard boxes or red/orange bags under any circumstances.</b></p>	<p><u>Infectious waste sharps include:</u> All needles, syringes, and blades; broken or unbroken glass and plasticware that has contacted infectious agents or was used in animal or human patient care or treatment, including plastic pipettes and other used plasticware that is recognizable after autoclaving or made of plastic that shatters on breakage or is considered breakable by the investigator.</p> <p><u>Chemical contaminated sharps waste:</u> Segregate from other sharps into purchased sharps containers labeled "Chemical Contaminated Sharps Waste-Do Not Autoclave"</p> <p><u>Non-infectious Sharps (needles, syringes, and blades):</u> Dispose in a sharps container through the infectious waste stream. Containers of non-infectious sharps may be discarded as infectious waste without prior autoclaving.</p> <p>NOTE: Non-infectious sharps may be commingled with infectious waste sharps or chemical-contaminated sharps waste. If this is done, they must be managed as described for the respective category of sharps.</p>	<p>All contaminated laboratory items <u>NOT</u> considered sharps, including used plasticware that is deformed after autoclaving or made of polyethylene (PE), polyvinyl chloride (PVC), and other thermoplastic polymers provided they do not shatter on breakage or are considered unbreakable by the investigator.</p>	<p>Infectious waste fluids (volumes greater than 20 cc).</p>	<p>Infectious waste that has been autoclaved.</p> <p><u>Chemical contaminated sharps containers</u> and <u>Non-infectious sharps containers (with label defaced)</u> that have NOT been autoclaved.</p>
<b>Reusable</b>	NO	<p><u>Reusable sharps containers</u> – YES</p> <p><u>Purchased sharps containers</u> – NO</p>	NO	YES	YES
<b>Disposal Method</b>	<p>Seal box when ¾ full for housekeepers to remove.</p>	<p><u>Reusable sharps containers:</u> vendor will remove full, closed containers and replace with empty ones</p> <p><u>Disposable sharps containers:</u> autoclave containers and discard in infectious waste bins.</p> <p><u>Chemical contaminated and Non-infectious Sharps Waste (with label defaced) :</u> discard <b>without</b> autoclaving into infectious waste bins.</p>	<p>Autoclave red/orange bagged infectious waste and discard in infectious waste bins.</p>	<p>Autoclave liquid or use appropriate disinfectant before pouring into a sanitary sewer drain.</p> <p>The reusable container must be washed and autoclaved before reuse.</p>	<p>Space Planning and Operations empties waste bins regularly.</p>
<b>Container Source</b>	User must purchase.	<p><u>Disposable sharps containers:</u> – user must purchase</p> <p><u>Reusable containers for sharps:</u> – delivered by vendor: Curtis Bay (Mr. Jim Frank @ 610-301-2913)</p>	User must purchase red/orange infectious waste autoclave bags.	User must purchase.	Space Planning and Operations (8-8000)

\*\* For more details on how to properly sort and manage disposal of ALL laboratory sharps waste, consult the EHRS LABORATORY SHARPS WASTE MANAGEMENT PROCEDURE (<http://www.ehrs.upenn.edu/programs/bio/waste/usedsharps.html>)